



DEPARTMENT OF HEALTH CARE, HUMAN SERVICE & FAMILY WELFARE  
GOVERNMENT OF SIKKM  
GANGTOK

Memo No:101/SHMW/2018-19

Dated:12/09/2018

**CORRIGENDUM**

The following corrigendum are hereby made by H.C, H.S & F.W. Department, Government of Sikkim on the Tender Notice for “Supply and installation of Blood Bank equipments for District Hospitals Singtam, Gyalshing, Mangan” vide Tender no. 02/SHMS//2018-19 as per advertisement R.O. no 291/IPR/PUB/Classi/18-19, dated 20/08/2018 and which was hosted at www.sikkim.gov.in.

1) **Section III** Page no. 7

**3.2 Important Dates**

**5** Bid Submission End Date: Postponed to **24/09/2018**, 1200 HRS IST.

**6** Bid Opening Date: Postponed to **24/09/2018**, 1300 HRS IST.

2) **Section VII Tender Details :**

**7.11.2 Valid Annual Turn Over statement:** Page no.14 (**This clause stands deleted**)

(The Bidder (Manufacturers/Authorized Dealers/representatives/Indian Subsidiaries/Direct Importers) should submit latest copy of **Valid Annual Turnover Statement**, verified by Chartered Accountant, showing the annual turnover of **Rs 200 lakhs** for the last three completed financial years.) **Stands Deleted.**

3) **Annexure I** Page no. 26-34

**Technical Specifications**

SL. NO	PARTICULARS
1	<b>Blood Bank refrigerator with alarm system (300 bags)-400 Lts</b> (Sl. nos. 5,6,19)
	5. The external cabinet should be rust proof material and have internal SS sheet and should have sliding trays made of Stainless steel (at least 4 trays for each Blood group). 6. Should have an inner plexi door for each compartment separately for easy viewing the blood bags. Outer lockable double glass door with air gaps. ( It should be: Inner Trays must be fitted with Plexi Glasses and Door should have an insulation of minimum 80mm with provision to view Inside of refrigerator from outside) 9. Internal temperature hold overtime in case of power failure should be at least 3-6 hrs.
2	<b>Refrigerator with Temp display for reagent/kits and samples (300 lts)</b> (Sl. nos. 1,3,6,13,14,15,16,18)
	1. Temperature: Between +2° C to +8° C 3. Display: 4 x 7 segment LED (Red)/ LCD 6. <b>Cabinet material: Ext. MS sheet (18 swg), and exterior powder coated after 7 tank process. Internal – Stainless steel sheet (22 swg)</b> (this line stands deleted) 13. Volume: 300ltr. 14. <b>Door: Single door MS exterior</b> (this line stands deleted) 15. <b>Ext. dimensions (WxDxH) mm: 600x630x1785</b> approximately (this line stands deleted) 16. <b>Approx. weight Kg.: 110</b> (this line stands deleted) 18 Tray: Plastic Coated trays (Adjustable coated wire mesh shelves)
3	<b>ELISA reader and washer with inbuilt printer</b> (Sl. no. 16)
	16.Should have inbuilt or external thermal printer for direct printing from the equipments 15 rolls of thermal papers should be supplied along with the unit.
5	<b>Gel Card Incubator (Dual Mode for Tube and Gel Card)</b> Spelling corrected

The above corrections may please be noted. Other conditions shall remain the same. For further clarifications if any, please contact the Office of the Superintending Engineer (Mech.), HC, HS & FW Deptt., Health Annexe Building, Convoy Ground, Tadong.

Superintending Engineer (Mech.),  
HC,HS & FW Deptt.,

