



CENTRAL HEALTH STORE ORGANIZATION
DEPARTMENT OF HEALTH CARE, HUMAN SERVICE & FAMILY WELFARE
GOVERNMENT OF SIKKM

Memo No: 294/CHSO

Dated: 04/08/2018

CORRIGENDUM

The following corrigendum are hereby made on the Tender document for "THE SUPPLY OF CT 100 MA X- RAY MACHINE AND ULTRASOUND MACHINE FOR VARIOUS HEALTH INSTITUTIONS FOR THE YEAR 2018-2019" as per advertisement memo no CHSO/239/2018, dated 14/07/2018.

1. **Non Conviction Certificate. (Clause 7, Page no 5):**

The Bidders (Manufacturers/ Authorized Dealers/ representatives/ Indian Subsidiaries /Direct Importers) should submit a latest copy of Notarized Affidavit Valid Non Conviction Certificate, as per format in Annexure IV in the technical bid, proving that the firms has not been blacklisted/ debarred by any Tender Inviting Authority or by any State Government or Central Government department/organization.

2. **Supporting Documents: (Clause 10, Page no 6):**

All pages of the bid shall be signed by the authorized person or persons signing the bid along with the stamp of the tenderer. The supporting documentary evidence submitted along with the Tender shall be produced duly attested by Gazatted Officer of the Government of Sikkim or Notarized. Any interlineations, erasures or over writing shall be valid only if they are initialed by the person (s) signing the offer. In the event of documentary proof as required being not enclosed, the Tender shall be liable to be rejected.

3. **Payment Terms: (Clause 24):**

The payment shall be made after full execution of the supply, installation and commissioning as ordered and submission of requisite documents and subject to other terms and conditions as laid down under the Terms and Conditions.

4. **Comparative statement of the technical specifications(Clause 27 to be added):**

The bidder should submit the Comparative statement of the technical specifications and compliance with the suppliers offered model, deviations and justifications.

5. **Technical Specifications: (Annexure VII):**

a) **Specifications of Computerized Radiography (CR) System**

Image Recording System (Cassettes & Image Plates) :

i) **Cassettes**

10" x 12"

- 2 nos. (to be added):

Image Reader (CR Reader) / Digitizer

- Processing Capacity 45 Plates/hour. (for 14" x 17" size- **stands deleted**)

Dry Laser Camera /Imager:

8. It should have two online tray, capable of changing any film size at any given point of time. Should support all films sizes such as 14x17, 14x14, 11x14, 10x12 & 8x10. (The changing of film size should be such that only film packet needs to be changed and not the whole tray to avoid wears and tear- **this line stands deleted**)
9. The system should be supplied with suitable UPS with a backup of 30 mins.
11. The system should be CE/ FDA approved.

b) TECHNICAL SPECIFICATIONS OF HIGH FREQUENCY MOBILE X-RAY MACHINE

A 100mA, 120KV high frequency X-Ray machine with mobile stand should be provided.

RATINGS

Power output – 4.2 KW upto 5KW
Rad KV - 40 to 120KV
Rad mA - 100mA or more
mAs -1- 200mAs

POWER SUPPLY

The quoted model should be AERB Certified./Type approved

c) Ultrasound Machine:

Point 2.1 Ultrasound Colour Doppler Machine with integrated trolley with probe, soft touch alphanumeric key board with track ball.

Point 3- 2 Nos Probes holder & 1 nos Gel holder conveniently to be placed.


Following transducers to be supplied:

No.2 **to be deleted.**

Point k – Frame rate minimum **500 FPS**, hard disk capacity of **500GB** or more.

6. BOQ (Annexure I):

Items	Qty required
1. Ultrasound machine	2 nos
2. 100 MA X Ray Machine	2 nos
3. CR system with printer	2 nos
4. Cost of film 8" X 10"	1 nos
5. Cost of film 10" X 12"	1 nos


Joint Director,
Central Health Stores Organization.
Central Health Stores Organization
HC, HS & F W. Department
Govt. of Sikkim, Gangtok